

ABSTRACT OF THE DISCLOSURE

The medical display includes a display device of a matrix type having a resolution of 100 to 300 ppi to display a medical image and at least one anti-reflection layer on a side of a front surface of said display device. The medical display system includes the medical display and a luminance meter measuring luminance. The anti-reflection layer has an average specular reflectivity of 0.5% or less at an incident angle of 5° in a wavelength range of 450 to 650 nm, receives light from a CIE standard light source D65 at an incident angle of 5° in a wavelength range of 380 to 780 nm to reflect the light as regular reflection light whose color falls within a range of $-7 \leq a^* \leq 7$ and $-10 \leq b^* \leq 10$ in terms of a^* and b^* values of CIE 1976 $L^*a^*b^*$ color space, and is placed on a surface whose flatness is defined by an arithmetic average height R_a and a maximum height R_z according to JIS B 0601-2001, with R_a set at 0.02 μm or less and R_z set at 0.04 μm or less.